


RAYCAST® HIGH PRECISION BASE PLATES HIGH PRECISION BASE PLATE IN CARBON FIBRE LAMINATE



Article No.: 32130/12

A. GENERAL PRODUCT INFORMATION

This product is a medical device used for patient positioning and immobilisation (head) in radiation therapy. This low density plate can be used during both the simulation and the treatment stage.

This product may only be used in combination with immobilisation masks produced by Orfit. Orfit prohibits the use of unauthorized third-party products in conjunction with its own products.

B. PRODUCT DESCRIPTION

This RAYCAST® HIGH PRECISION BASE PLATE in Carbon Fibre Laminate (CFL) is suitable for use in combination with the range of Efficast® thermoplastic Head & Neck masks, Raycast® High Precision Head Supports, Blocks & Wedges and Couch Indexing Devices to form a reproducible patient positioning and immobilisation device in the field of radiotherapy.

Information on these other parts and instructions on how to make the masks can be found in the respective 'instructions for use' and on www.orfit.com.

C. PRECAUTIONS FOR USE

Always place the base plate on a flat surface. Clear the treatment table of any debris before positioning the base plate and make sure that it is completely supported by the couch.

Always fix the base plate securely to the treatment couch before positioning a patient. More information on how to fix this base

plate to a couch top can be found on our website http://www.orfit.com/en/fixation-devices/.

The CFL base plate is constructed to be light in weight and to have superb dosimetric properties. When handled roughly, it may get damaged and fibres may come off. When this happens, stop using the base plate to prevent fibres from getting in contact with the patient's or user's skin. Contact your distributor.

A patient set-up form is available on the Orfit website: https://www.orfit.com/app/uploads/Head-and-Neck.pdf

D. STORAGE

Always store the product in a safe place to prevent it from getting damaged or falling onto other objects. Take care not to damage the edges of the plate when storing it in an upright position. Prevent hard objects from falling onto the plate. Store the base plate between +10°C (50°F) and 30°C (86°F).

E. PROPERTIES

E.1 Physical Properties

This plate is made of a carbon fibre laminate.

Dimensions: L 287 mm x W 195 mm x H 12 mm


L 11.29" x W 7.67" x 0.47"

Weight: 0.23 kg

0.51 lbs

E.2 Dosimetric Properties

| Position | Attenuation | Attenuation | Skin build-up |
|----------|-------------|-------------|---------------|
| | (± 0.15 %) | (± 0.15 %) | (± 0.1 mm) |
| | (6MV) | (15MV) | (15MV) |
| | % | % | mm H₂O |
| | | | equiv. |
| 1 | 1.07 | 0.55 | 2.94 |
| 2 | 1.65 | 0.97 | 4.79 |



Note: Use these numbers as a guidance only. Perform the measurements again in your department to verify these results. Position 2 refers to all locations contouring slots, holes etc.

F. MAINTENANCE AND WASTE MANAGEMENT

This product can be cleaned and disinfected by means of soapy water or an isopropanol based disinfectant, applied with a soft cloth. If unsure about the cleaning fluid, do not use. **Never use aerosol sprays, corrosive cleaning agents, solvents or abrasive detergents.**

Periodic checks of the product should be done to insure the parts are not worn and require repair or replacement. **Do not attempt to make repairs yourself.** Contact your distributor if there are any questions or concerns.

The product can be disposed of with household waste.

G. ADDITIONAL INFORMATION

For additional information such as distributor contact information, product brochures, Safety Data Sheets and regulatory information, please visit our website www.orfit.com.

Note:

It is prohibited to make alterations to this text without prior approval from Orfit Industries. $RAYCAST^* \ and \ EFFICAST^* \ are \ registered \ trademarks of Orfit Industries.$



ORFIT INDUSTRIES

Vosveld 9A | B-2110 Wijnegem | Belgium T (+32) (0)3 326 20 26

welcome@orfit.com

www.orfit.com

Ref. No. 50183 VERSION 1 LAST UPDATE: 13/04/2021

REVISION DATE: 13/04/2021